
Patient-Specific Drill Guide Template for Pedicle Screw Insertion into the Atlantoaxial Cervical Spine Utilizing Stereolithographic Modelling: An In Vitro Study

Harjoland L. Obenieta, MD, Rafael C. Bundoc, MD
Department of Orthopedics, University of the Philippines Manila, Philippine General Hospital, Philippines

ABSTRACT

Introduction. Pedicle screw fixation is a recognized treatment option for cervical instabilities. However, the use of cervical pedicle screw fixation is limited by its demand on technical skill and risks owing of injury to the neurovascular structures. Hence, the accuracy of screw placement is vital in the success of the procedure. In the recent decade, attention has been given to 3D printing to create drill guide templates in cervical spine stabilization. While various literature has described the accuracy of using patient-specific drill guide in the subaxial cervical spine, limited studies explored on its role in the atlantoaxial spine. **Objectives.** To assess the accuracy of atlantoaxial cervical pedicle screw insertion using patient-specific drill guide template from stereolithographic modelling. **Design.** In vitro (cadaveric) study **Methods**: Seventeen atlantoaxial cervical vertebrae specimens were collected from the University of the Philippines Manila, College of Medicine, Department of Anatomy. Out of 17, two C1 and two C2 vertebrae were excluded. 3D reconstruction of cervical models was done using CT scan images of the specimens. These images were formatted to .stl file and were subsequently used to produce cervical plastic models. Using acrylic cement, drill guide templates were molded. Pedicle screw insertion was done on cadaveric specimens using the specific drill guide templates. The accuracy of screw placement was evaluated by an independent evaluator. **Results.** A total of 60 pedicles (combined C1 and C2) from 15 cadaveric axial cervical vertebrae were evaluated. Pedicle screw insertion in the antlantoaxial cervical vertebrae has a total accuracy of 93.33%. One-hundred percent accuracy was achieved in the C1 vertebrae compared to 86% in the C2 vertebrae. **Conclusion.** Patient-specific drill guide template using stereolithographic modelling is accurate in the pedicle screw insertion of cadaveric atlantoaxial specimens.

IMAGES



Overview of the steps taken in screw insertion. A, awl is used to make a hole following the trajectory of the drill guide template. B, Manual drilling. C, Tapping. D, Screw insertion

BIBLIOGRAPHY

1. Bundoc R, Grozman S, and Delgado G. (2016). A Novel Patient-Specific Drill Template for Pedicle Screw Insertion of the Subaxial Cervical Spine Using Stereolithography: An In Vitro Study. Unpublished manuscript, Department of Orthopedics, Philippine General Hospital, University of the Philippines Manila
2. Ianotti J, Baker J, Rodriguez E, Brems J, Ricchetti E, Mesiha M and Bryan J. (2014) Three-dimensional preoperative planning software and a novel information transfer technology improve glenoid component positioning. J Bone Joint Surg Am. 96:e71(1-8)
3. Pu X, Luo C, Lu T, Yao S, and Chen Q (2018). Clinical Application of Atlantoaxial Pedicle Screw Placement Assisted by a Modified 3D-Printed Navigation Template. Clinics 73:e259
4. Pu X, Luo C, Lu T, Yao S, and Chen Q (2018). Design and Application of a Novel Patient-Specific Three-Dimensional Printed Drill Navigational Guiding in Atlantoaxial Pedicle Screw Placement. World Neurosurg 114:e1-e10
5. Chachan S, Abd Razak HR, Loo WL, Allen JC, Kumar DS. Cervical pedicle screw instrumentation is more reliable with O-arm-based 3D navigation: analysis of cervical pedicle screw placement accuracy with O-arm-based 3D navigation. European Spine Journal. 2018 Nov 1;27(11):2729-36.



Dr. Harjoland Obenieta is currently a 3rd year orthopedic surgery resident in the University of the Philippines Manila, Philippine General Hospital. He has particular interest in spine surgery and research.

E-mail: hlobenieta@up.edu.ph