

Assessment of adhesion response to 3D printed materials for ophthalmic device development

ALBAND M [1], LEE RMH [1,2], PENNY M [1], BROCCHINI S [1,2], HILTON S[1]

[1]*UCL School of Pharmacy, 29-39 Brunswick Square, London, WC1N 1AX*
[2]*UCL Institute of Ophthalmology, 11-43 Bath Street, London, EC1V 9EL*

Introduction: Glaucoma is the leading cause of irreversible visual impairment worldwide. Glaucoma surgical devices fail due to a scarring response that results in fibrous encapsulation surrounding the device preventing aqueous humour drainage.

3D printing technology has the potential to develop personalised ophthalmic devices or organs with improved cost effectiveness and productivity. Limited experimental data exists as to the biocompatibility response of 3D printed photopolymers. We performed cell adhesion and protein adsorption studies of 3D printed photopolymers compared to materials used in current ophthalmic devices (Silicone, Polytetrafluoroethylene (PTFE) and Poly (methyl methacrylate) (PMMA)) to assess 3D printed materials as a potential route for ophthalmic device development.

Methods: 3D printed materials (n=6) were developed using a high-resolution, desktop stereo-lithography (SLA) 3D printer and compared to materials used in current ophthalmic devices. Protein adsorption was quantified using a micro bicinchoninic acid (Micro BCA) assay and fluorescein-conjugated bovine serum albumin (FITC-BSA) adsorption. Cell adhesion (monocytes, fibroblasts) was assessed using Alamar Blue, CyQUANT and Live/Dead assays. Data were compared using a two-tailed unpaired t-Test.

Results: 3D printed materials demonstrated low cell adhesion and protein adsorption. Results were similar to those found with materials used in current ophthalmic devices ($P>0.05$). However it was noted that 3D printed materials demonstrated increased cytotoxicity ($P<0.05$).

Conclusion: 3D printed photopolymer materials demonstrated a similar biocompatibility response to currently used materials and may allow for the development of customisable ophthalmic devices or organs. Subsequent testing will determine the adhesion response to 3D printed materials containing anti-scarring agents.

Biography

Maryam Alband completed her MPharm degree with first class honours in 2013 at UCL School of Pharmacy, London. Following her registration as a fully qualified pharmacist at Royal Free NHS Foundation Trust, she returned to UCL to undertake a PhD in Translational Sciences and Pharmaceutics. Her work involved the development of a novel medical implant funded by a prestigious NIHR grant that has led to the award of an international travel grant.

Presenting author details
Full name: Maryam Alband

Contact number: +44 (0) 757 286 0952
Twitter account: n/a
Linked In account: n/a
Session name/ number:
Category: Poster